

BESA License Key 3.0 - Intended Use and Product Classification

Documents are subject to workflows, and content shall only be used if the content is approved (indicated by a green ring-symbol)

🟢 Approved, Changed on: February 22nd 2024 at 9:12:46 AM (UTC)



Intended Use and Product Classification

Purpose:

This document contains information concerning the intended use and the product classification of BESA License Key 3.0 in English language.

Product Name

BESA License Key 3.0

Intended Use

The BESA License Key 3.0 provides copy protection for BESA software and can protect one or more BESA programs. Individual BESA products or modules can be activated using HASP features (slots) within the BESA License Key 3.0.

- The purpose of the BESA License Key 3.0 is to provide the licensing for BESA products.
- The BESA License Key 3.0 is only to be used in accordance with the intended use.
- The license key is allowed to be used as an accessory product for the following BESA products:
 - BESA Research (7.1 January 2024)
 - BESA MRI (4.0)
- The BESA License Key 3.0 is intended to be used:
 - Locally – attached to the PC running the BESA program (see above), or
 - Remotely within a LAN

Please note that the BESA License Key 3.0 needs to be attached to the machine or the network throughout the entire session of the relevant BESA program.

Please note that in a remote access scenario, the required drivers for using the BESA License Key 3.0 (available for installation as "HASPUsetsetup.exe") need to be installed locally on the machine running the BESA program, and on the machine to which the BESA License Key 3.0 is attached.

- The BESA License Key 3.0 is not intended to be used over a VPN tunnel or other remote LAN access methods.
- The BESA License Key 3.0 is not for commercial use, such as fee-based analysis services.

Specific Medical Purposes

BESA License Key has no specific medical purpose.

Intended User

The intended users are:

- Administrator:
 - Installation of software required by the BESA License Key 3.0.
 - Configuration of network environment (if applicable).
 - Installation of license updates provided by BESA GmbH.
- User of BESA programs:
 - Installation of license updates provided by BESA GmbH.

Please note, that beside potential installation of license updates, if initial configuration of BESA License Key 3.0 (and network environment, if not used locally) has been done correctly, no action from user of BESA products with respect to BESA License Key 3.0 is required.

BESA GmbH assumes no liability for unauthorized access to, or unauthorized use of the product.

The product shall not be used by children, laypersons or patients.

Product Classification

According to Regulation (EU) 2017/745, BESA License Key 3.0 is not a medical device in itself, but to be classified as an accessory to BESA medical devices.

Medical Device Class	
BESA License Key is an accessory for a medical device	according to MDR - EU Regulation 2017/745; Article 2
Designed for long term use	according to MDR - EU Regulation 2017/745; ANNEX VIII: CLASSIFICATION RULES; CHAPTER I: DEFINITIONS SPECIFIC TO CLASSIFICATION RULES; DURATION OF USE; Section 1.3
Class: no class assigned, according to classification BESA License Key is not a medical device.	according to MDR - EU Regulation 2017/745, Article 2: Definitions

Software Safety Classification	
BESA License Key is not a medical device, therefore no safety classification is required	N/A

Page ID

242811020

Change History (valid revision number)

Version	Effective Date	Approvals	Version Notes
1	Thu, Feb 22, 2024, 09:10:13 AM	>> *Action: approve, MK, Approval: QM Approval (State: QM Approval, Thu, Feb 22, 2024, 09:10:13 AM)	None
2	Thu, Feb 22, 2024, 09:12:46 AM	>> *Action: approve, MK, Approval: QM Approval (State: QM Approval, Thu, Feb 22, 2024, 09:12:46 AM)	None

Template Change History

TEMPLATE	TEMPLATE Version	Effective Date	Version Notes
Intended Use and Product Classification	1	Tue, 14 Mar 2023 08:35:55 GMT	None
Intended Use and Product Classification	2	Tue, 02 May 2023 08:21:48 GMT	None
Intended Use and Product Classification	3	Mon, 12 Jun 2023 11:26:33 GMT	None
Intended Use and Product Classification	4	Mon, 26 Jun 2023 14:40:42 GMT	None

Printed version is NO controlled document and NOT subject to modifications